

PRODUCT SPECIFICATIONS

FOR

POWDER FREE NITRILE EXAMINATION GLOVE 3.2MIL - SE

Material: Nitrile Latex and compounding materials safe for use in medical and food handling.
Description: Ambidextrous, Beaded Cuff, Micro-textured, Chlorinated, Powder Free
Available Colour: Orange

Standards of Conformance:

Product conforms to the following test standards

EN 455-1	Medical gloves for single use – Part 1: Requirements and testing freedom from holes
EN 455-2	Medical gloves for single use – Part 2: Requirements and testing for physical properties
EN 455-3	Medical gloves for single use – Part 3: Requirements and testing for biological evaluation
ISO 2859-1	Sampling Procedures and Tables for Inspection by Attributes

Performance Requirements					
Characteristic	Inspection Level			AQL	
Watertight Test	G1			1.5	
Visual Test	S4			Major 2.5 Minor 4.0	
Physical Dimension	N = 13			Median value obtained must comply	
Physical Properties	N = 13			Median value obtained must comply	
Particulate Residue	N = 5			≤ 2 mg/glove	
Physical Dimensions					
SIZE	Extra-Small	Small	Medium	Large	Extra-Large
Median Palm Width ± 4 mm	76	86	98	107	115
Median Glove Length (mm)	Minimum 240				
Median Thickness ± 0.02 mm					
Palm (center of palm)	0.08				
Finger (13 ± 3 mm from tip)	0.12				
Physical Properties					
Test	Before Aging			After Aging	
Median Force at Break	Minimum 6N			Minimum 6N	

November 2014